


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVIR-2.-C-P.

S.N.: 8951

Date: Aug 11 1999



- a) Output signal amplitude:
0 to +200 V (40V)
- b) Pulse width: (50ns)
- c) Rise time:
2 to 10 ns
- d) Fall time:
1/2 ns
- e) PRF: 0 to 30 kHz
- f) Jitter, stability:

- g) Prime power:
OK
120/240V
50-60Hz.

60 db ATTEN

100 VOLT/DIV

PRF ≈ 10 kHz