


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AN-1010-C*

S.N.: *8954*

Date: *JUNE 4 1999*



- a) Output signal amplitude: *0 TO ±100 VOLTS ✓*
- b) Pulse width: *(R_L = 50Ω) 50 NS TO 10MS ✓*
- c) Rise time: *(10% MAX DUTY CYCLE) ≤ 15NS ✓*
- d) Fall time: *≤ 15NS ✓*
- e) PRF: *0 TO 1.0 MHz ✓*
(10% MAX DUTY CYCLE)
- f) Jitter, stability: *OK ✓*
- g) Prime power: *120/240 V₀ 50 to 1MHz*



(B) *Next 30 V RANGE*
R_L = 50Ω 10V/DIV
500 NS/DIV