


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVX-FD3-PS .

S.N.: 8955

Date: July 5, 1999



- a) Output signal amplitude: ECL to 50Ω to ground or -2V ✓
- b) Division Factor: 2 to 65535 ✓
- c)
- d)
- e) PRF: 0 - 250MHz ✓
- f) Jitter, stability: OK ✓
- g) Prime power: 120/240V_{AC}, 50-60Hz ✓

ECL OUT 1
Outputs for

- 1) N=1
- 2) N=2
- 3) N=9

$F_{IN} = 250\text{MHz}$

$R_L = 50\Omega$ to ground