


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVL-2-C-PN

S.N.: 8965

Date: Aug 20 1999



a) Output signal amplitude:
0 to 5350 V (50 Ω)

b) Pulse width:
5 to 100 ns

c) Rise time:
 ≤ 2 ns

d) Fall time:
 ≤ 3 ns

e) PRF:
0 to 5 kHz

f) Jitter, stability: OK

g) Prime power: 120/240 V
50-60 Hz

60 db ATTEN
 ≈ 100 V/DIV
PRF ≈ 2 kHz
10 ns/DIV