


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVL-AV-1-P-P5-T*

S.N.: *3967*

Date: *AUG 13 1999*



- a) Output signal amplitude:
0 TO + 100 VOLTS
- b) Pulse width:
3 TO 100 NS
- c) Rise time:
1 TO 10 NS
- d) Fall time:
2 TO 10 NS
- e) PRF:
0 TO 5 KHz

- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz

≈ 60 dB ATTEN
≈ 100 VOLTS/DIV
TOP 10 NS/DIV

T CONTROL AT MIN,
MAX & TWO MID VALUES

Bot 1 us/DIV
T CONTROL AT MIN
PRF ≈ 3 KHz