


PULSE GENERATOR
PERFORMANCE CHECK

Model: AV-156A-C-P-OP1B-ISIA (repair)

S.N.: 8974

Date: May 14, 2003



a) Output Signal Amplitude: 0-5A

b) Pulse Width: 1 - 100 ms

c) Rise Time: < 10 μ s

d) Fall Time: < 10 μ s

e) PRF: 1 - 100 Hz

f) Jitter, Stability: OK

g) Prime Power: 120/240V AC, 50-60 Hz,

h) Inter-channel delay: 0 - 1000 ms

Top: OUT1, 1.75A to 5.018 Ω , 2ms PW, 100 ms delay
Bottom: OUT2, 1.75A to 5.018 Ω , 2ms PW
Both: Manual triggering, 5V / div, 20 ms / div.