


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVM-1-C-A-EW-EO-EO-P*

S.N.: *8983*

Date: *Oct 6 1999*



- a) Output signal amplitude:
0 to +5 Volts (50Ω)
- b) Pulse width:
0.2 to 6 ns
- c) Rise time:
≤ 100 ps
- d) Fall time:
≤ 135 ps
- e) PRF: *0 to 25 MHz*
- f) Jitter, stability:
OK
- g) Prime power: *120/200V*

40 dB ATTEN
0.5 Volts/Div
TOP 1 ns/Div, PRF ≈ 3 MHz
MID AS TOP BUT 100ps/Div
BOT 5 ns/Div, PRF ≈ 25 MHz

50-60 Hz.