

PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVO-7B2-C-P-WA*

S.N.: *8984*

Date: *SEPT 20 1999*

- a) Output signal amplitude:  
*0 TO + 225 VOLTS*
- b) Pulse width:  
*R<sub>L</sub> > 4 OHMS (50 AMP MAX)*  
*1 TO 5 US*
- c) Rise time:  
*≤ 0.5US*
- d) Fall time:  
*≤ 0.5US*
- e) PRF: *0 TO 100 Hz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V*  
*50-60Hz*

SN 8934  $R_L = 50 \Omega$

