

PULSE GENERATOR
PERFORMANCE CHECK

Model: *APR-5B-B-N-LUB*

S.N.: *3939*

Date: *SEPT 30 1999*

- a) Output signal amplitude:
0 TO -500 VOLTS
- b) Pulse width:
(TO $R_L > 1 \text{ M}\Omega$)
10 μ S TO 2 SEC
(50% MAX DUTY CYCLE)
- c) Rise time:
 $\leq 200 \text{ NS}$
- d) Fall time:
 $\leq 200 \text{ NS}$
- e) PRF:
0.1 MHz TO 1 kHz
(50% MAX DUTY CYCLE)
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz

[Signature]





