


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-2-CN*

S.N.: *8990*

Date: *April 24 1999*



- a) Output signal amplitude: *0 TO -15 VOLTS*
- b) Pulse width: *0.2 TO 2.0 NS*
- c) Rise time: *≤ 100 ps*
- d) Fall time: *≤ 135 ps*
- e) PRF: *0 TO 25 MHz*

f) Jitter, stability: *OK*

g) Prime power: *120/240V*
50-60 Hz
AJ

40dB ATTEN

-10 VOLTS / DIV

TOP 1 NS / DIV

PRF ≈ 10 MHz

BT 5 NS / DIV

PRF ≈ 25 MHz