


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVO-8.C-C-P-EW*

S.N.: *9002*

Date: *OCT 26 1999*



- a) Output signal amplitude:
0 TO +40V TO
- b) Pulse width:
R_L 70.2 ~ (200 AMPS IN PR. 2µS TO 20ms (+DC))

c) Rise time:
≤ 1µS

d) Fall time:
≤ 1µS


e) PRF: *0 TO 1KHz*

f) Jitter, stability: *OK*

g) Prime power: *a) 120 / 240V
50-60 Hz*

*b) 0 TO +40 Volts,
100 AMPS DC.*

*ⓐ LOAD VOLTAGE, R_L
= 0.2 ~ WITH 12"
LONG LEADS. 50 AMP/DIV*



ⓑ AS ⓐ BUT PW ~ 10ms