


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-5-1-C-T

S.N.: 9004

Date: AUG 16 1999



a) Output signal amplitude:
0 to +10V (50Ω)

b) Pulse width:
≤ 130 ps.

c) Rise time:
NA

d) Fall time:
NA

e) PRF: 0 to 1 MHz

f) Jitter, stability: OK

g) Prime power: 120/240V
50-60 Hz

40 db ATTEN
- 10 VOLTS/DIV

DP0 2 ns/DIV

MID 500 ps/DIV

BOT 100 ps/DIV

PRF ≈ 200 kHz