


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AV-107C-B-N*

S.N.: *9005*

Date: *SEPT 13 1999*



a) Output signal amplitude:
0 TO -10 AMP TO
b) Pulse width:
0 TO 60 VOLTS
100 NS TO 1.0 US

c) Rise time: *≤ 20NS*

d) Fall time: *≤ 20NS*

e) PRF: *0 TO 5 KHz*

f) Jitter, stability:

g) Prime power: *OK*
120/240V
50-60 Hz.

R_L = 4.7 Ω

TIP: MONITOR OUT
TO 50 Ω

BOT: LOAD VOLTAGE