

PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV-151B-C*


S.N.: *9009*

Date: *OCT 20 1999*


- a) Output signal amplitude:
0 to $\pm 200V$ ($R_L \geq 50K$)
- b) Pulse width:
*SINE, SQUARE PULSES
& DC OFFSET*
- c) Rise time:
 $\leq 1 \mu s$
- d) Fall time:
 $\leq 1 \mu s$
- e) PRF:
0 to 100 kHz
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V
50-60Hz*
- h) GAIN: *0 to X100*
- i) DC: *0 to $\pm 50VDC$*

[Signature]

$$R_c = 50K.$$



$$h_c = 50K$$



$$R_L = 50K$$

