


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV-110B-PS*
S.N.: *9018*
Date: *NOV 9 1999*

- a) Output signal amplitude:
0 TO ± 200 V ($R_L > 50\Omega$)
- b) GAIN :
 $\times 1$ TO $\times 100$
- c) Rise time:
 $\leq 1\mu s$ ($V_0 - P_0$)
- d) Fall time:
 $\leq 1\mu s$
- e) PRF:
0 TO 100 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz



9010
 $R_c = 50k$

V_{out}

V_{in}

