


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVPP-1-C-N*


S.N.: *9020*

Date: *SEPT 23 1999*



- a) Output signal amplitude: *0 TO -10V (TO 50Ω)*
- b) Pulse width: *1) 0.2 TO 5.0 NS*
2) 5.0 TO 100 NS
- c) Rise time: *≤ 100ps*
- d) Fall time: *≤ 135ps*
- e) PRF: *0 TO 1.0 MHz*

ⓐ *PG A OUT.*
40dB ATTN. PRF ≈ 100KHz



1NS/DIV
10NS/DIV

- f) Jitter, stability: *OK*
- g) Prime power: *120/240V*
50-60Hz.

ⓑ *PG B OUT*
40dB ATTN.
PRF ≈ 100KHz