


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVIR-1-PS-FOLCA-P-ECL-EVA-EW

S.N.: 9021

Date: NOV 15 1999



- a) Output signal amplitude: 0 TO +200 VOLTS
- b) Pulse width: 1 TO 5 NS
- c) Rise time: ≤ 0.5 NS
- d) Fall time: ≤ 1 NS
- e) PRF: 0 TO 50 KHz.
- f) Jitter, stability: OK
- g) Prime power: 120/240V

60 db ATTN
100 VOLTS/DIV
TR PW NORMAL MAX
BOT PW NORMAL MIN.
PRF = 30 KHz

50-60Hz