


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVIR-1-PS-FOICAT-FCI-EA-CW

S.N.: 9022

Date: NOV 15 1999



- a) Output signal amplitude: 0 to +200 Volts
- b) Pulse width: (to 50ns)
- c) Rise time: 1 to 5 ns
≤ 0.5 ns
- d) Fall time: ≤ 10 ns
- e) PRF: 0 to 50 kHz
- f) Jitter, stability: OK
- g) Prime power: 120/240V
50-60 Hz

60 db ATTEN

100 VOLTS/DIV

PW ≈ MAX

AND PW ≈ MIN.

PRF = 30 KHz

[Signature]