


PULSE GENERATOR

PERFORMANCE CHECK

Model: *MHF-1*  
S.N.: *9030*  
Date: *NOV 8 1999*



- a) Output signal amplitude: *> 10 VOLTS*
- b) Pulse width: *≤ 130 ps*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *100 ± 5 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *0.5 WATT (15 Vp-p), 100 MHz CW*

*40 dB ATTN*  
*10 VOLTS/DIV*  
*TOP 2 ns/DIV*  
*BOT 200 ps/DIV*

*[Signature]*