


PULSE GENERATOR

PERFORMANCE CHECK

Model: *ANMP-2-P*

S.N.: *9038*

Date: *SEPT 29 1999*



- a) Output signal amplitude: \checkmark
0 TO +10 VOLTS
- b) Pulse width: *(50ns)*
- c) Rise time: *5 TO 100 NS*
- d) Fall time: $\leq 100ps$
- e) PRF: $\leq 135pp$
- f) Jitter, stability: \checkmark
OK
- g) Prime power: \checkmark
+ 24 VOLTS
300mA MAX

40 db ATTN
10 VOLTS/DIV
TOP 10 NS/DIV
BOT 200ps/DIV (RISE TIME)