


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *ANMP-2-P*

S.N.: *9040*

Date: *SEPT 29 1999*



- a) Output signal amplitude: ✓  
*0 to +10 VOLTS*
- b) Pulse width: *(50ns)*
- c) Rise time: *5 to 100 ns* ✓
- d) Fall time: *≤ 100 ps* ✓
- e) PRF: *≤ 135 ps* ✓  
*0 to 1.0 MHz*
- f) Jitter, stability: ✓  
*OK*
- g) Prime power:

*+ 24 VOLTS*  
*300mA MAX*

*40 dB ATTN*  
*10 VOLTS/DIV*  
*TOP 20 ns/DIV*  
*BOT 200 ps/DIV (RISE TIME)*