


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVR-E2-C-P-05

S.N.: 9044

Date: JAN 18 2000



- a) Output signal amplitude: 0 TO +50 VOLT (7050mV)
- b) Pulse width: 20 TO 200 NS
- c) Rise time: ≤ 0.5 NS
- d) Fall time: ≤ 0.5 NS
- e) PRF: 0 TO 50 KHz
- f) Jitter, stability: OK
- g) Prime power: 120/240V

50-60 Hz

50 db ATTEN

32 VOLTS/DIV

TDR 10 NS/DIV

MMUD 1.0 NS/DIV (RISE TIME)

BoT 1.0 NS/DIV (FALL TIME)

PRF \approx 50 KHz