


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ANMR-2-PT-M1*

S.N.: *9048*

Date: *NOV 29 1999*



- a) Output signal amplitude: *0 TO +10 V (50%)*
- b) Pulse width: *20 TO 200NS*
- c) Rise time: *(10% MAX DUTY CYCLE) ≤ 300ps*
- d) Fall time: *≤ 300ps*
- e) PRF: *0 TO 1.0 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *+24VDC*
300mA MAX.
- h) DEOFFSE: *0 TO ±5 Volts*

40 db ATTEN
-10 V/20ns/10V

TOP 20 ns/DIV
TRIG 1 NS/DIV (RISE TIME)
TRIG 1 NS/DIV (FALL TIME)
PAUSE = 300 kHz