


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVL-2-CN

S.N.: 9049

Date: NOV 17 1999



- a) Output signal amplitude: 0 to -350V (to 50Ω)
- b) Pulse width: 5 to 100ns
- c) Rise time: $\leq 2ns$
- d) Fall time: $\leq 3ns$
- e) PRF: 0 to 5KHz
- f) Jitter, stability: 0%
- g) Prime power: 120/240V

60 db ATTEN
TR 10ns/DIV
P505 2ns/DIV
(LEAD IN EP 60)

50-60Hz
[Signature]