

PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVR-G1-C-P*

S.N.: *9052*


Date: *NOV 11 1999*

- a) Output signal amplitude:  
*0 to +250V (RL > 10K)*
- b) Pulse width:  
*100 NS TO 1 MS*  
*(20% MAX DUTY CYCLE)*
- c) Rise time:  
*≤ 10 NS*
- d) Fall time:  
*≤ 20 NS*
- e) PRF: *0 TO 1 KHz*  
*(20% MAX DUTY CYCLE)*
- f) Jitter, stability:  
*OK*
- g) Prime power: *120/240V*  
*50-60Hz.*

*[Signature]*

9052  
 $R_c = 10K.$

$V_{out}$



9052.

$R_L = 10k.$

RISE TIME.

$V_{out}$

