


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AV14-1M2-P*

S.N.: *9058*

Date: *SEPT 20 1999*



a) Output signal amplitude:

$\approx \pm 60 \text{ V (50}\Omega\text{)}$

b) Pulse width:

$\leq 2 \text{ ns}$

c) Rise time:

NA

d) Fall time:

NA

e) PRF:

0 TO 10 MHz

f) Jitter, stability:

OK

g) Prime power:

$\pm 135 \pm 5 \text{ VDC}$

3 mA max

50 db ATTN

$\approx 32 \text{ VOLTS/DIV}$

PRF $\approx 300 \text{ kHz}$