

PULSE GENERATOR
PERFORMANCE CHECK

Model: *AC 110B-PS*
S.N.: *9074*
Date: *NOV 29 1999*

- a) Output signal amplitude:
0 TO ± 200 V ($R_L \geq 50K$)
- b) ~~Gain~~:
X 1 TO X 100
- c) Rise time:
 $\leq 1 \mu s$
- d) Fall time:
 $\leq 1 \mu s$
- e) PRF: *0 TO 100 KHz.*
- f) Jitter, stability: *OK.*
- g) Prime power: *120/240V*
50-60 Hz.

A.J.

9074

V_{out}

V_{in}

