


PULSE GENERATOR

PERFORMANCE CHECK

Model: AV6-3B-C P-M

S.N.: 9078

Date: DEC 16 1999



- a) Output signal amplitude:
0 to +450 Volts (70.5mV)
- b) Pulse width:
 $\leq 2 \text{ ns}$
- c) Rise time:
 $\leq 0.8 \text{ ns}$
- d) Fall time:
 $\leq 0.8 \text{ ns}$
- e) PRF:
0 to 20 KHz
- f) Jitter, stability:
OK
- g) Prime power: 120/245V

60 db ATTN
~ 100 Volts / DIV
PRF ~ 5 KHz

50-60 KHz

[Handwritten signature]