

PULSE GENERATOR
PERFORMANCE CHECK



Model: *AVR-4-PS-PN*

S.N.: *9082*

Date: *DEC 14 1999*

- a) Output signal amplitude: *0 TO ±400 VOLTS*
- b) Pulse width: *($R_L \geq 50\Omega$)
0.1 TO 100 μ S*
- c) Rise time: *(0.5% MAX DUTY CYCLE)
 ≤ 20 NS*
- d) Fall time: *≤ 20 NS*
- e) PRF: *0 TO 10 KHZ
(0.5% MAX DUTY CYCLE)*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240 V
50-60 HZ.*

Q082
NARROW PULSE.
 $P_{ic} = 50 \mu$
P OUT.



9082
WIDE PULSE
 $R_L = 50 \Omega$
N out.

