

PULSE GENERATOR

PERFORMANCE CHECK


Model: *AV-110B-PS-LTA*

S.N.: *9087*

Date: *NOV 4 1999*

- a) Output signal amplitude: *0 TO ± 200 VOLTS*
- b) GAIN: ($R_L \geq 50K$)
X1 TO X100
- c) Rise time: *≤ 1 μs (V_{0-P})*
- d) Fall time: *≤ 1 μs*
- e) PRF: *0 TO 100 KHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240 V*
50-60 Hz
- h) OFFSET: *0 TO ± 150 Volts*

9087
 $R_L = 50K.$



9087
 $R_c = 50K$

OUT

IN

