

PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV-110B-PS-LTA*

S.N.: *9088*

Date: *NOV 4 1999*


- a) Output signal amplitude:
0 TO ± 200 VOLTS
- b) GAIN :: ($R_L \geq 50K$)
X1 TO X100
- c) Rise time:
≤ 1 μS (V_{0-p})
- d) Fall time:
≤ 1 μS
- e) PRF:
0 TO 100 KHz
- f) Jitter, stability:
OK.
- g) Prime power: *120/240V*
50-60Hz
- h) OFFSET: *0 TO ± 150*
VOLTS



9033
 $R_c = 50k$

V_{out}

V_{in}



9038
 $R_L = 50K$

V_{out}

V_{in}

