

PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AV-110B-R5-LTA*

S.N.: *9089*

Date: *NOV 8 1999*


- a) Output signal amplitude:  
*0 TO  $\pm 200$  Volts ( $R_L \geq 50k$ )*
- b) GAIN :  
*X1 TO X100*
- c) Rise time:  
 *$\leq 145$  (V<sub>0-1</sub>)*
- d) Fall time:  
 *$\leq 145$*
- e) PRF:  
*0 TO 100 kHz*
- f) Jitter, stability:  
*OK*
- g) Prime power: *170/240 V*  
*50-60 Hz*
- h) OFFSET: *0 TO  $\pm 150$  V*



90 29  
 $R_c = 50k$

$V_{out}$

$V_{in}$



9089  
 $R_C = 50K.$

