

PULSE GENERATOR
PERFORMANCE CHECK

Model: PW-110B-PS-LTA

S.N.: 9090

Date: NOV 8 1999

- a) Output signal amplitude:
0 TO ± 200 VOLTS
- b) GAIN : $R_L \geq 50\Omega$
 $\times 1$ TO $\times 100$
- c) Rise time:
 ≤ 145 (V_{0-P})
- d) Fall time:
 ≤ 145
- e) PRF:
0 TO 100 KHz.
- f) Jitter, stability:
OK
- g) Prime power: 120/240 V
50-60 Hz
- h) OFFSET: 0 TO ± 150 V.




4

9090
 $R_L = 50K$

V_{out}

V_{in}



4

9090

$R_L = 50K$

V_{out}

V_{in}

