


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVH-MPT-EBG1D*

S.N.: *9092*

Date: *FEB 7 2000*



- a) Output signal amplitude: *> +20 VOLTS (TO 500)*
- b) Pulse width: *600 ps (FWHM)*
- c) Rise time: *≤ 150 ps*
- d) Fall time: *≤ 275 ps*
- e) PRF: *0 TO 2 MHz.*
- f) Jitter, stability: *OK*

- g) Prime power: *+15 VOLTS*
250 mA MAX.

40 dB ATTN

10 VOLTS / DIV

TOP: 500 ps / DIV

BOT: 100 ps / DIV (RISE TIME)

PRF ≈ 100 MHz

[Signature]