


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AN4-MPT-EB61D*

S.N.: *9093*

Date: *FEB 7 2000*



- a) Output signal amplitude:  
 *$\geq +200.48$  ( $70.50 \Omega$ )*
- b) Pulse width:  
*600 ns (FWHM)*
- c) Rise time:  
 *$\leq 150 \mu s$*
- d) Fall time:  
 *$\leq 275 \mu s$*
- e) PRF:  
*0 TO 2.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*+15 VDC*  
*2.50 mA MAX*

*40 dB ATTEN*

*10 V OUTS / DIV*

*TRF: 500 ps / DIV*

*BOF: 100 ps / DIV (RISE TIME)*

*PRF = 1.0 MHz*