

PULSE GENERATOR

PERFORMANCE CHECK

Model: AV-110B-PS-LTA

S.N.: 9106

Date: NOV 23 1999

a) Output signal amplitude:

0 TO ± 200 VOLTS ($R_L \geq 75\Omega$)

b) GAIN: $X / TO \times 100$

c) Rise time:

$\leq 1\mu S$ (V_{0.7})

d) Fall time:

$\leq 1\mu S$

e) PRF:

0 TO 100 KHz

f) Jitter, stability:

OK

g) Prime power:

120/240V

50-60 Hz


h) OFFSET: 0 TO ± 150 VOLTS

06

9106
 $R_c = 50K.$


V_{out}

V_{in}



9106
 $R_L = 50K$

V_{out}
 V_{in}



24 Nov 1999
12:48:33