


PULSE GENERATOR
PERFORMANCE CHECK

Model: AP-M-1-P

S.N.: 9111

Date: JAN 25 2000



- a) Output signal amplitude:
0 TO +10 VOLTS (50Ω)
- b) Pulse width:
0.2 TO 4.0 NS
- c) Rise time:
≤ 100 ps
- d) Fall time:
≤ 135 ps
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 VDC

300 mA (MAX)

40 dB ATTEN
10 V0 25Ω / DIV
TDP 1.0 ns / DIV
V0I 200 ps / DIV
PRF = 100 kHz