


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVHF-3

S.N.: 9120

Date: DEC 21 1999



- a) Output signal amplitude: ^{*}
-10V
- b) Pulse width:
≤ 130ps
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
500 MHz
- f) Jitter, stability:
OK
- g) Prime power: 500 MHz
0.5 WATT

40dB ATTEN
-10 VOLT/DIV
TOP: 500ps/DIV
BOT: 200ps/DIV

* 15 VOLTS/UP
[Signature]

* 15V SPEC IN TABLE
ON PAGE 78 OF CAT 10
IS A MISPRINT.