


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR-5B-B-7N*

S.N.: *9123*

Date: *DEC 10 1999*

- a) Output signal amplitude:
0 TO ± 500 VOLTS
- b) Pulse width: ($R_L \geq 50 \Omega$)
0.1 TO 100 μ S
- c) Rise time: (*1% MAX DUTY CYCLE*)
≤ 30N)
- d) Fall time:
≤ 30NS
- e) PRF: *0 TO 10 KHZ*
(10% MAX DUTY CYCLE)
- f) Jitter, stability:
OK
- g) Prime power: *120/240V*
50-60 Hz





9123

$R_i = 50\Omega$

N out

"NARROW PULSE"



9123

$R_L = 50 \Omega$

P_{out}

" WIDE PULSE "

