


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVP-AU-14V3-C-P*

S.N.: *9125*

Date: *JAN 17 2000*



- a) Output signal amplitude:  
*0 TO +40 Volts*
- b) Pulse width:  
*(to 50Ω)*  
*0.4 TO 2.0 NS*
- c) Rise time:  
*≤ 150ps*
- d) Fall time:  
*≤ 250ps*
- e) PRF:  
*0 TO 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*100 ± 10 Watts*  
*50-60 Hz*

*50 dB ATTEN*  
*32 Volts/DIV*

*[Signature]*