


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ANL-2-C-N*
S.N.: *9129.*
Date: *DEC 3 1999*



60 db ATTN
100 V/83/DIV
TOP 10 NS/DIV
Bot 2 NS/DIV
(RISE TIME)

PRF ≈ 2 KHz

- a) Output signal amplitude:
0 TO -350V (TO 50M)
- b) Pulse width:
≤ TO 100 NS
- c) Rise time:
≤ 2 NS
- d) Fall time:
≤ 3 NS
- e) PRF: *0 TO 5 KHz*
- f) Jitter, stability:
OK
- g) Prime power: *120/240V*
50-60 Hz