


PULSE GENERATOR

PERFORMANCE CHECK

Model: *ARM-1-P*

S.N.: *9143*

Date: *JAN 27 2000*



- a) Output signal amplitude:
0 TO +5 VOLTS (50Ω)
- b) Pulse width:
0.2 TO 6.0 NS
- c) Rise time:
≤ 100 ps
- d) Fall time:
≤ 135 ps
- e) PRF:
0 TO 25 MHz.
- f) Jitter, stability:
OK
- g) Prime power:
+24 VDC
500mA MAX

40dB ATTN

5 VOLTS/DIV

TRIP 5 NS/DIV (≈ 25 MHz)

PROT 1 NS/DIV (≈ 1 MHz)