


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVPP-1-C-P*  
 S.N.: *9155*  
 Date: *FEB 29 2000*



- a) Output signal amplitude: *0 TO +10V (TO 50Ω)*
- ① b) Pulse width: *A) 0.2 TO 5.0 NS*  
*B) 5.0 TO 100 NS*
- ② c) Rise time: *≤ 100 ps*
- ③ d) Fall time: *≤ 135 ps*
- ④ e) PRF: *0 TO 1.0 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V*

*40 db ATTEN.: 10 VOLTS/DIV*

*50-60 Hz.*

- ① A out *1 NS/DIV*
- ② A out *100 ps/DIV (RISE TIME)*
- ③ B out *5 NS/DIV*
- ④ B out *100 ps/DIV (RISE TIME)*

*PRF = 100 kHz.*

*[Signature]*