


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVI-MP-N*
 S.N.: *9178*
 Date: *MAR 10 2000*



- a) Output signal amplitude:
≥ 40V (50Ω)
- b) Pulse width:
2 TO 100 NS
- c) Rise time:
≤ 350 ps
- d) Fall time:
≤ 2 NS
- e) PRF:
PW = 2 NS : 0 TO 1.0 MHz
PW = 100 NS : 0 TO 20 kHz
- f) Jitter, stability:
OK
- g) Prime power: *+15 VDC*
150 mA MAX

[Signature]

50 dbb ATTEN
≈ 32 VOLTS/DIV
TOP DL ≈ 24"
BOT DL ≈ 0"
PRF ≈ 100 kHz