


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *A20-7B2-B-F*  
S.N.: *9187*  
Date: *MAR 23 2000*

- a) Output signal amplitude:  
*0 to + 150 VOLTS*
- b) Pulse width: *(R<sub>L</sub> > 1.0Ω)*  
*1 to 10 μS*
- c) Rise time:  
*≤ 0.5 μS*
- d) Fall time:  
*≤ 0.5 μS*
- e) PRF: *0 to 100 Hz.*
- f) Jitter, stability: *OK*
- g) Prime power:  
*120/240V*  
*50-60 Hz*



$R_L = 1.0 \Omega$   
PRF = 10 MHz  
LOAD VOLTAGE

