

PULSE GENERATOR
PERFORMANCE CHECK

Model: *M-110B-PS-LTA*

S.N.: *9208*

Date: *APRIL 2 2000*

- a) Output signal amplitude:
0 TO ± 200 VOLTS (R.P. 50A)
- b) GAIN:
× 1 TO × 100
- c) Rise time:
≤ 145 (V_{0-P})
- d) Fall time:
≤ 145
- e) PRF:
0 TO 100 KHz
- f) Jitter, stability:
OK
- g) Prime power: *120/240V*
50-60 Hz.
- h) OFFSET: *0 TO ± 150V*



9200
 $R_c = 50k$

V_{out}

V_{in}

