


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-2-C-P-D*

S.N.: *9214*

Date: *APRIL 10 2000*



- a) Output signal amplitude: *0 TO +15 VOLTS (50 OHMS)*
- b) Pulse width: *0.2 TO 2.0 NS*
- c) Rise time: *≤ 100 ps*
- d) Fall time: *≤ 135 ps*
- e) PRF: *0 TO 25 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240 V*
50 - 60 MHz

40 dB ATTEN
10 VOLTS/DIV
TOP 5 NS/DIV
BOI 200 ps/DIV
PRF = 25 MHz