


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVHF-1

S.N.: 9216

Date: APRIL 10 2000



- a) Output signal amplitude: <sup>\*</sup>  
≈ 10 VOLTS
- b) Pulse width:  
≤ 130 ps
- c) Rise time:  
NA
- d) Fall time:  
NA
- e) PRF: 100 ± 5 MHz
- f) Jitter, stability: OK
- g) Prime power:  
0.5 WATT (15 V<sub>p-p</sub>)  
100 MHz CW

*[Handwritten signature]*

40 db ATTEN  
10 VOLTS/DIV  
TR 2ns/DIV  
BO.T 200ps/DIV

\* 15V SPEC IN TABLE ON PAGE 78 OF CAT 10 IS A MISPRINT