


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVHF-3*

S.N.: *9217*

Date: *APRIL 10 2000*



- a) Output signal amplitude: *10 V ✓*
- b) Pulse width: *≤ 130 ps ✓*
- c) Rise time: *NA ✓*
- d) Fall time: *NA ✓*
- e) PRF: *500 MHz ✓*
- f) Jitter, stability: *OK ✓*
- g) Prime power: *500 MHz ✓*
0.5 WATT ✓

40 dB ATTEN
10 VOLT/DIV
TOP: 500 ps/DIV
BOT: 200 ps/DIV

15 VOLT/DIV
15 VOLT/DIV
15 V SPEC IN TABLE
ON PAGE 78 OF CAT 10
17 A MISPRINT.