


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVIMP-2-C-PEPIA*

S.N.: *9222*

Date: *FEB 15 2000*



- a) Output signal amplitude: *✓*
0 TO +10 V (70.50 Ω)
- b) Pulse width:
5 TO 100 NS \checkmark
- c) Rise time:
 ≤ 100 ps \checkmark
- d) Fall time:
 ≤ 135 ps \checkmark
- e) PRF:
0 TO 100 MHz \checkmark
- f) Jitter, stability:
OK \checkmark
- g) Prime power:
120/240 V
50-60
WY

40 dB ATTEN

-- 10 V/DIV

TOP 5 NS/DIV

MID 200ps/DIV (RISE TIME)

BOT 200ps/DIV (FALL TIME)

PRF \approx 100 kHz